

Certificate of Analysis

Lab Ref No.: **PL2011-C37205T1**

SN: **RS2875275875148936**

Page 1 / 2



Ooi Kah Wai

Chemist

IKM No.: L/2452/7352/16

AN1: RA6393022301069526

PAC Meditech Sdn Bhd

NO. 760 – 2, JALAN CASSIA SELATAN 3/5, TAMAN PERINDUSTRIAN

BATU KAWAN,

14110 BANDAR CASSIA PULAU PINANG

MALAYSIA

Date of Issue:

2020-11-17

Date of Received:

2020-11-06

Date of Completion:

2020-11-17

Sample Marking:

Face Mask

Splash Resistance Pressure

Specimen	Splash Resistance Pressue (kPa)	Status
1	16	Pass
2	16	Pass
3	16	Pass
4	16	Pass
5	16	Pass
6	16	Pass
7	16	Pass
8	16	Pass
9	16	Pass
10	16	Pass
11	16	Pass
12	16	Pass
13	16	Pass
14	16	Pass
15	16	Pass
16	16	Pass
17	16	Pass
18	16	Pass
19	16	Pass
20	16	Pass
21	16	Pass
22	16	Pass
23	16	Pass



Certificate of Analysis

Lab Ref No.: **PL2011-C37205T1**

SN: **RS2875275875148936**

Page 2 / 2



Ooi Kah Wai

Chemist

IKM No.: L/2452/7352/16

AN1: RA6393022301069526

Specimen	Splash Resistance Pressure (kPa)	Status
24	16	Pass
25	16	Pass
26	16	Pass
27	16	Pass
28	16	Pass
29	16	Pass
30	16	Pass
31	16	Pass
32	16	Pass

Remark: The specification refer to Face Mask EN14683:2019 Type IIR.

Test Description	Unit	Result(s)	Spec	Method or Equipment Used
Splash Resistance Pressure*	kPa	Pass at 16	≥ 16.0	ISO 22609:2004

– END OF REPORT –

This Test Report is prepared according to our standard testing requirements and based solely on the samples submitted by you. We have made every reasonable effort to ensure the accuracy and reliability of the information provided herein. However, we shall not be deemed to have warranted or guaranteed the accuracy and reliability of the results contained in the Test Report under whatsoever circumstances. Neither shall we be liable and responsible for any loss or damage of whatever nature (direct, indirect, incidental, economical, punitive and consequential) caused or alleged to be caused by or in connection with and arising out of the use of or reliance on this Test Report.

